
Midmark M11 - v1.0.5
Total Cycles: 3729

Name: RITTER2
Sterilizer ID: V2313228
06/21/2024 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.2 F 0.0
2:00 151.3 F 2.5
4:00 165.4 F 6.0
6:00 192.0 F 10.7
8:00 212.9 F 4.8
10:00 225.8 F 9.8
12:00 231.4 F 9.0
14:00 240.1 F 12.3
16:00 240.0 F 11.0
18:00 252.8 F 17.5
20:00 264.2 F 24.2
22:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.7 F 29.9
2:00 271.5 F 29.2
3:00 271.5 F 29.7
4:00 272.6 F 29.8

Min 271.1 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 211.7 F 0.0
10:00 208.6 F 0.0
15:00 205.9 F 0.0
20:00 207.7 F 0.0
25:00 211.0 F 0.0
30:00 208.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

06/21/2024 03:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____