

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2756

Name: RITTER2  
Sterilizer ID: V2313228  
08/14/2023 07:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.9 F 0.0  
2:00 71.4 F 0.0  
4:00 106.9 F 0.0  
6:00 124.1 F 0.0  
8:00 152.9 F 0.0  
10:00 178.0 F 0.0  
12:00 201.0 F 0.9  
14:00 211.4 F 4.0  
16:00 217.4 F 6.8  
18:00 226.3 F 9.8  
20:00 234.9 F 13.3  
22:00 233.6 F 9.0  
24:00 243.0 F 13.5  
26:00 241.1 F 11.3  
28:00 238.2 F 9.8  
30:00 251.7 F 16.4  
32:00 262.9 F 22.9  
34:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.2 F 28.9  
2:00 272.7 F 29.4  
3:00 271.8 F 29.0  
4:00 272.0 F 28.8

Min 271.5 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 218.1 F 0.0  
5:00 216.6 F 0.0  
10:00 220.9 F 0.0  
15:00 224.0 F 0.0  
20:00 221.6 F 0.0  
25:00 222.4 F 0.0  
30:00 221.3 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 34:15  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:12:17  
CYCLE COMPLETE

08/14/2023 09:00 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_