

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1231

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2021 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.7 F 0.0  
2:00 96.6 F 0.0  
4:00 112.8 F 0.0  
6:00 156.8 F 0.1  
8:00 182.0 F 0.2  
10:00 202.9 F 0.8  
12:00 214.8 F 3.9  
14:00 223.6 F 6.7  
16:00 230.7 F 10.0  
18:00 238.3 F 11.1  
20:00 236.4 F 9.8  
22:00 241.5 F 10.0  
24:00 245.3 F 12.9  
26:00 243.2 F 11.6  
28:00 255.5 F 17.8  
30:00 266.1 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 272.6 F 28.6  
2:00 271.8 F 27.9  
3:00 272.1 F 28.4  
4:00 272.5 F 28.4

Min 271.4 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 198.5 F 0.0  
10:00 200.8 F 0.0  
15:00 201.5 F 0.0  
20:00 202.7 F 0.0  
25:00 202.3 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 31:43  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:02  
CYCLE COMPLETE

10/05/2021 10:22 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01