
Midmark M11 - v1.0.5
Total Cycles: 958

Name: RITTER2
Sterilizer ID: V2313228
04/05/2022 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 186.3 F 0.4
4:00 205.0 F 4.2
6:00 222.4 F 9.4
8:00 232.2 F 8.7
10:00 237.1 F 11.5
12:00 237.4 F 10.0
14:00 238.4 F 9.2
16:00 251.9 F 16.6
18:00 264.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 272.0 F 28.9
3:00 271.7 F 29.2
4:00 272.7 F 29.6

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 212.5 F 0.0
10:00 223.2 F 0.0
15:00 231.1 F 0.0
20:00 224.5 F 0.0
25:00 219.8 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

04/05/2022 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____