
Midmark M11 - v1.0.5
Total Cycles: 160

Name: RITTER2
Sterilizer ID: V2313228
07/05/2021 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 162.3 F 2.7
4:00 175.4 F 6.5
6:00 195.7 F 11.4
8:00 215.7 F 6.4
10:00 230.0 F 11.6
12:00 232.1 F 8.9
14:00 238.7 F 10.6
16:00 245.7 F 13.8
18:00 257.7 F 20.3
20:00 268.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.0
2:00 271.6 F 29.4
3:00 271.7 F 29.9
4:00 272.4 F 29.9

Min 271.1 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 223.3 F 0.0
10:00 221.3 F 0.0
15:00 220.7 F 0.0
20:00 212.4 F 0.0
25:00 201.6 F 0.0
30:00 196.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

07/05/2021 12:28 PM
ACCEPTED: X
REJECTED:
Operator: U01