
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006835
12:45 17/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:02
STERILIZING 3:08
136.4 C 224kPa 3:08
136.1 C 222kPa 3:38
136.2 C 221kPa 4:08
136.4 C 223kPa 4:38
136.3 C 222kPa 5:08
136.0 C 222kPa 5:38
136.2 C 223kPa 6:08
136.4 C 224kPa 6:38
136.4 C 224kPa 6:38
VENTING 6:39
AIR DRYING 7:26
CYCLE COMPLETE 7:32

Digital Signature #
B133A79D150DA9B9
