
Midmark M11 - v1.0.0
Total Cycles: 3902

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 144.7 F 2.4
4:00 162.8 F 5.9
6:00 195.3 F 10.2
8:00 211.3 F 4.6
10:00 223.8 F 8.3
12:00 236.5 F 13.1
14:00 235.1 F 9.6
16:00 239.4 F 9.3
18:00 246.3 F 13.5
20:00 258.2 F 19.9
22:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.5 F 29.2
2:00 271.5 F 28.4
3:00 272.6 F 29.2
4:00 271.8 F 28.6

Min 271.2 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 223.9 F 0.0
10:00 224.9 F 0.0
15:00 237.4 F 0.0
20:00 218.3 F 0.0
25:00 213.7 F 0.0
30:00 210.6 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

10/13/2023 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____