
Midmark M11 - v1.0.5
Total Cycles: 276

Name: RITTER2
Sterilizer ID: V2313228
08/10/2021 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 159.6 F 1.7
4:00 164.9 F 5.4
6:00 193.0 F 10.0
8:00 209.0 F 6.2
10:00 224.2 F 10.3
12:00 227.7 F 8.1
14:00 238.6 F 11.9
16:00 240.8 F 12.9
18:00 256.8 F 19.3
20:00 267.7 F 25.9
22:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.5 F 29.6
2:00 272.6 F 30.0
3:00 271.4 F 29.4
4:00 273.3 F 29.9

Min 271.2 F 29.3
Max 273.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 229.7 F 0.0
10:00 228.8 F 0.0
15:00 232.2 F 0.0
20:00 218.3 F 0.0
25:00 209.8 F 0.0
30:00 173.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

08/10/2021 11:14 AM
ACCEPTED: X
REJECTED:
Operator: U01