
Midmark M11 - v1.0.5
Total Cycles: 3862

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.9 F 0.0
2:00 138.9 F 1.9
4:00 168.0 F 5.1
6:00 191.6 F 9.1
8:00 204.6 F 11.2
10:00 218.5 F 6.7
12:00 230.3 F 11.3
14:00 231.0 F 8.0
16:00 243.7 F 12.8
18:00 244.0 F 12.6
20:00 256.1 F 19.1
22:00 266.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 29.4
2:00 272.1 F 29.6
3:00 272.3 F 29.2
4:00 272.2 F 29.7

Min 271.5 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.3 F 0.0
5:00 205.4 F 0.0
10:00 202.3 F 0.0
15:00 200.0 F 0.0
20:00 196.6 F 0.0
25:00 194.2 F 0.0
30:00 191.4 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 23:28
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:29
CYCLE COMPLETE

08/14/2024 06:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____