

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4211

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2024 11:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 95.6 F 0.0  
2:00 110.9 F 0.7  
4:00 145.3 F 2.6  
6:00 152.1 F 6.0  
8:00 175.1 F 9.8  
10:00 205.3 F 11.0  
12:00 212.6 F 5.7  
14:00 225.9 F 9.8  
16:00 235.4 F 9.3  
18:00 236.8 F 10.4  
20:00 237.4 F 8.0  
22:00 246.8 F 13.9  
24:00 258.3 F 20.1  
26:00 268.8 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.2  
2:00 271.5 F 28.4  
3:00 272.4 F 29.3  
4:00 272.1 F 28.8

Min 271.2 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 212.8 F 0.0  
10:00 211.9 F 0.0  
15:00 217.3 F 0.0  
20:00 219.6 F 0.0  
25:00 230.1 F 0.0  
30:00 231.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 27:12  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:05:39  
CYCLE COMPLETE

01/17/2024 12:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_