
Midmark M11 - v1.0.5
Total Cycles: 74

Name: RITTER2
Sterilizer ID: V2313228
06/08/2021 06:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 73.2 F 0.2
4:00 98.0 F 1.1
6:00 117.2 F 3.6
8:00 141.5 F 6.8
10:00 161.2 F 11.0
12:00 201.0 F 3.1
14:00 208.3 F 7.6
16:00 220.9 F 11.9
18:00 228.9 F 8.1
20:00 242.3 F 13.5
22:00 241.8 F 11.9
24:00 253.8 F 18.2
26:00 265.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.9
2:00 272.1 F 30.0
3:00 272.2 F 29.9
4:00 271.9 F 29.7

Min 270.6 F 28.9
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 209.5 F 0.0
10:00 215.7 F 0.0
15:00 217.4 F 0.0
20:00 221.1 F 0.0
25:00 223.1 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 27:48
STERILIZING: 4:00
VENTING: 2:51
DRYING: 30:00
TOTAL CYCLE: 01:05:30
CYCLE COMPLETE

06/08/2021 07:47 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____