
Midmark M11 - v1.0.0
Total Cycles: 4883

Name: RITTER 1
Sterilizer ID: V2178198
09/10/2024 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 88.7 F 0.0
4:00 92.1 F 0.0
6:00 117.4 F 0.1
8:00 147.3 F 0.4
10:00 178.8 F 0.8
12:00 197.6 F 3.4
14:00 212.7 F 6.5
16:00 223.1 F 10.0
18:00 230.1 F 10.7
20:00 232.8 F 8.7
22:00 243.7 F 13.5
24:00 241.1 F 10.9
26:00 236.9 F 8.8
28:00 250.9 F 15.4
30:00 262.2 F 21.8
32:00 272.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.7
2:00 271.6 F 27.8
3:00 272.5 F 28.7
4:00 272.4 F 28.4

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 213.6 F 0.0
10:00 225.2 F 0.0
15:00 234.5 F 0.0
20:00 231.6 F 0.0
25:00 214.6 F 0.0
30:00 214.1 F 0.0

Drying Complete

FILLING: 0:41
HEATING: 32:29
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:32
CYCLE COMPLETE

09/10/2024 09:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____