

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2274

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2023 09:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
2:00 176.8 F 3.1  
4:00 196.2 F 8.1  
6:00 213.3 F 9.8  
8:00 229.2 F 9.5  
10:00 234.8 F 8.9  
12:00 242.4 F 11.9  
14:00 246.4 F 13.7  
16:00 259.9 F 21.0  
18:00 271.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.2 F 29.1  
2:00 271.5 F 28.9  
3:00 272.5 F 29.2  
4:00 271.3 F 28.6  
  
Min 271.3 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 227.2 F 0.0  
10:00 227.5 F 0.0  
15:00 228.1 F 0.0  
20:00 229.1 F 0.0  
25:00 228.0 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 1:53  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:39  
CYCLE COMPLETE

04/20/2023 10:28 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_