
Midmark M11 - v1.0.5
Total Cycles: 2755

Name: RITTER2
Sterilizer ID: V2313228
08/11/2023 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 149.6 F 1.4
4:00 158.4 F 4.6
6:00 187.6 F 9.7
8:00 210.7 F 4.7
10:00 231.7 F 10.4
12:00 233.4 F 8.4
14:00 238.7 F 9.7
16:00 248.5 F 14.8
18:00 250.9 F 16.0
20:00 265.1 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.3 F 28.8
2:00 272.9 F 29.4
3:00 271.8 F 28.9
4:00 272.4 F 28.9

Min 271.6 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 226.1 F 0.0
5:00 216.5 F 0.0
10:00 214.9 F 0.0
15:00 215.7 F 0.0
20:00 207.2 F 0.0
25:00 203.0 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:37
STERILIZING: 4:00
VENTING: 2:53
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

08/11/2023 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____