

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005456  
7:53 02/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:05  
STERILIZING 4:09  
136.4 C 223kPa 4:09  
136.4 C 221kPa 4:39  
136.0 C 222kPa 5:09  
136.4 C 222kPa 5:39  
136.4 C 222kPa 6:09  
136.4 C 222kPa 6:39  
136.1 C 217kPa 7:09  
136.4 C 223kPa 7:39  
136.4 C 223kPa 7:39  
VENTING 7:39  
AIR DRYING 8:27  
CYCLE COMPLETE 12:12

-----  
Digital Signature #  
F9063D25A6794FF3  
-----