
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004684
10:53 08/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:44
PRESSURIZING 1:51
STERILIZING 2:58
136.4 C 223kPa 2:58
136.4 C 224kPa 3:28
136.4 C 224kPa 3:58
136.4 C 224kPa 4:28
136.3 C 222kPa 4:58
136.4 C 224kPa 5:28
136.4 C 224kPa 5:58
136.0 C 217kPa 6:28
136.0 C 217kPa 6:28
VENTING 6:28
AIR DRYING 7:19
CYCLE COMPLETE 9:21

Digital Signature #
07AD3E2C858CA4AB
