

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3963

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2024 12:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.2 F 0.0  
2:00 138.0 F 1.6  
4:00 154.5 F 4.1  
6:00 178.9 F 7.4  
8:00 200.8 F 11.5  
10:00 213.6 F 4.5  
12:00 220.9 F 8.4  
14:00 233.5 F 12.6  
16:00 234.0 F 9.5  
18:00 238.7 F 10.4  
20:00 245.0 F 13.4  
22:00 256.1 F 19.3  
24:00 266.2 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.4 F 30.1  
2:00 271.8 F 29.5  
3:00 271.2 F 29.5  
4:00 272.5 F 30.1

Min 271.1 F 29.2  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 204.3 F 0.0  
10:00 202.4 F 0.0  
15:00 202.1 F 0.0  
20:00 197.2 F 0.0  
25:00 194.9 F 0.0  
30:00 193.8 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 25:48  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:04:55  
CYCLE COMPLETE

09/20/2024 01:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_