

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3788

Name: RITTER 1  
Sterilizer ID: V2178198  
09/18/2023 11:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 174.8 F 2.3  
4:00 191.1 F 6.4  
6:00 209.1 F 10.9  
8:00 225.8 F 5.3  
10:00 229.8 F 10.7  
12:00 232.6 F 7.5  
14:00 243.4 F 13.2  
16:00 242.6 F 11.9  
18:00 255.0 F 18.2  
20:00 266.0 F 24.8  
22:00 272.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 271.5 F 28.8  
2:00 271.6 F 28.9  
3:00 271.5 F 28.8  
4:00 271.5 F 28.7

Min 271.5 F 28.7  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.8 F 0.0  
5:00 207.4 F 0.0  
10:00 208.4 F 0.0  
15:00 198.1 F 0.0  
20:00 193.5 F 0.0  
25:00 194.5 F 0.0  
30:00 195.0 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 22:21  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:01:23  
CYCLE COMPLETE

09/18/2023 12:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_