
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002405
13:55 20/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:01
STERILIZING 3:04
136.0 C 220kPa 3:04
136.3 C 223kPa 3:34
136.3 C 225kPa 4:04
136.4 C 225kPa 4:34
136.4 C 225kPa 5:04
136.2 C 224kPa 5:34
136.2 C 223kPa 6:04
136.4 C 223kPa 6:34
136.4 C 223kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 16:30

Digital Signature #
A07F1F77E87AADDE
