

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002079  
9:50 23/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:45  
STERILIZING 3:59  
136.2 C 223kPa 3:59  
136.4 C 226kPa 4:29  
136.3 C 225kPa 4:59  
136.4 C 226kPa 5:29  
136.1 C 223kPa 5:59  
136.2 C 224kPa 6:29  
136.4 C 227kPa 6:59  
136.4 C 227kPa 7:29  
136.4 C 227kPa 7:29  
VENTING 7:29  
AIR DRYING 8:24  
CYCLE COMPLETE 8:46  
-----

Digital Signature #  
A5651EA1EB676DD0  
-----