
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001841
7:58 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:41
STERILIZING 4:55
136.0 C 221kPa 4:55
136.4 C 225kPa 5:25
136.3 C 223kPa 5:55
136.0 C 221kPa 6:25
136.4 C 224kPa 6:55
136.4 C 225kPa 7:25
136.3 C 224kPa 7:55
136.4 C 224kPa 8:25
136.4 C 224kPa 8:25
VENTING 8:25
AIR DRYING 9:17
CYCLE COMPLETE 9:27

Digital Signature #
25D205A06C2C4D19
