
Midmark M11 - v1.0.0
Total Cycles: 4281

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2024 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.4 F 0.0
2:00 172.1 F 3.2
4:00 185.6 F 7.6
6:00 207.1 F 12.7
8:00 221.6 F 6.6
10:00 233.8 F 11.7
12:00 234.3 F 8.9
14:00 239.1 F 9.8
16:00 247.7 F 14.0
18:00 259.9 F 20.7
20:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 28.8
2:00 271.1 F 28.2
3:00 272.6 F 29.0
4:00 271.8 F 28.4

Min 271.0 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 220.1 F 0.0
10:00 227.0 F 0.0
15:00 227.9 F 0.0
20:00 226.5 F 0.0
25:00 225.5 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

02/07/2024 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____