

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3324

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2023 03:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.6 F 0.0  
2:00 168.1 F 2.3  
4:00 181.4 F 6.3  
6:00 207.7 F 10.8  
8:00 220.4 F 4.9  
10:00 232.5 F 10.1  
12:00 236.3 F 8.1  
14:00 243.3 F 13.0  
16:00 242.4 F 11.6  
18:00 255.4 F 18.3  
20:00 266.9 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.7 F 29.3  
2:00 271.9 F 28.6  
3:00 271.7 F 28.9  
4:00 272.6 F 29.2  
  
Min 271.3 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 211.1 F 0.0  
10:00 216.4 F 0.0  
15:00 229.6 F 0.0  
20:00 230.3 F 0.0  
25:00 226.7 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 21:31  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:59:52  
CYCLE COMPLETE

06/01/2023 04:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_