
Midmark M11 - v1.0.0
Total Cycles: 2863

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2023 04:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.1 F 0.0
2:00 140.8 F 2.6
4:00 162.8 F 6.0
6:00 200.8 F 10.5
8:00 213.0 F 3.9
10:00 226.6 F 8.8
12:00 239.2 F 11.7
14:00 238.3 F 10.6
16:00 237.3 F 8.3
18:00 249.8 F 15.2
20:00 261.6 F 21.9
22:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.8
2:00 271.9 F 29.0
3:00 271.8 F 28.6
4:00 272.0 F 29.1

Min 271.3 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 230.7 F 0.0
10:00 243.7 F 0.0
15:00 235.6 F 0.0
20:00 232.5 F 0.0
25:00 225.4 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

02/06/2023 05:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____