
Midmark M11 - v1.0.5
Total Cycles: 1381

Name: RITTER2
Sterilizer ID: V2313228
08/09/2022 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 193.5 F 0.2
4:00 195.6 F 3.1
6:00 212.2 F 8.5
8:00 220.1 F 11.6
10:00 230.6 F 9.6
12:00 237.3 F 9.3
14:00 243.7 F 13.5
16:00 243.4 F 12.7
18:00 255.9 F 19.4
20:00 267.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.8 F 29.9
2:00 271.8 F 29.1
3:00 272.9 F 30.1
4:00 272.5 F 29.5

Min 271.7 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 211.8 F 0.0
10:00 220.3 F 0.0
15:00 229.8 F 0.0
20:00 230.7 F 0.0
25:00 225.9 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

08/09/2022 09:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____