
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001899
13:49 28/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:12
STERILIZING 4:13
136.0 C 220kPa 4:13
136.2 C 222kPa 4:43
136.3 C 223kPa 5:13
136.2 C 223kPa 5:43
136.2 C 224kPa 6:13
136.1 C 219kPa 6:43
136.4 C 225kPa 7:13
136.4 C 223kPa 7:43
136.4 C 223kPa 7:43
VENTING 7:43
AIR DRYING 8:39
CYCLE COMPLETE 8:44

Digital Signature #
EC20CF4EBCC7B127
