
Midmark M11 - v1.0.5
Total Cycles: 3571

Name: RITTER2
Sterilizer ID: V2313228
04/30/2024 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.8 F 0.0
2:00 135.9 F 1.3
4:00 149.8 F 4.1
6:00 172.4 F 7.5
8:00 191.2 F 11.4
10:00 208.5 F 4.0
12:00 219.6 F 8.0
14:00 230.7 F 12.3
16:00 232.3 F 8.7
18:00 242.3 F 11.8
20:00 243.7 F 12.7
22:00 255.2 F 18.7
24:00 265.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 30.0
2:00 272.7 F 29.5
3:00 271.6 F 29.0
4:00 272.0 F 29.6

Min 271.5 F 28.9
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 211.5 F 0.0
10:00 213.5 F 0.0
15:00 201.2 F 0.0
20:00 209.9 F 0.0
25:00 212.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:57
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:04:27
CYCLE COMPLETE

04/30/2024 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____