
Midmark M11 - v1.0.5
Total Cycles: 4317

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 173.2 F 1.7
4:00 169.7 F 5.0
6:00 182.8 F 9.2
8:00 202.7 F 10.6
10:00 222.0 F 7.2
12:00 231.6 F 11.8
14:00 231.8 F 8.6
16:00 242.9 F 12.1
18:00 243.9 F 12.9
20:00 255.4 F 19.1
22:00 265.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 271.9 F 30.2
2:00 272.9 F 30.2
3:00 272.0 F 29.8
4:00 271.2 F 29.4

Min 271.0 F 29.3
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 180.5 F 0.0
10:00 200.9 F 0.0
15:00 195.8 F 0.0
20:00 190.2 F 0.0
25:00 189.5 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

01/30/2025 12:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____