
Midmark M11 - v1.0.0
Total Cycles: 2297

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2022 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 184.1 F 1.0
4:00 208.9 F 4.1
6:00 211.2 F 8.7
8:00 224.0 F 10.8
10:00 230.1 F 9.1
12:00 238.3 F 10.6
14:00 242.0 F 11.9
16:00 240.0 F 10.4
18:00 253.5 F 17.0
20:00 265.0 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.3 F 29.0
2:00 271.9 F 28.3
3:00 272.1 F 28.8
4:00 272.2 F 28.6

Min 271.5 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 218.2 F 0.0
10:00 229.4 F 0.0
15:00 232.2 F 0.0
20:00 233.2 F 0.0
25:00 216.3 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

09/07/2022 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____