

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3583

Name: RITTER2  
Sterilizer ID: V2313228  
05/02/2024 01:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.6 F 0.0  
2:00 128.8 F 1.6  
4:00 147.2 F 4.3  
6:00 174.9 F 7.6  
8:00 192.6 F 11.5  
10:00 211.6 F 4.2  
12:00 221.2 F 8.6  
14:00 231.0 F 13.3  
16:00 235.2 F 10.3  
18:00 235.8 F 8.9  
20:00 246.5 F 14.3  
22:00 257.3 F 20.1  
24:00 266.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.3  
1:00 272.6 F 30.1  
2:00 272.2 F 29.6  
3:00 271.1 F 29.2  
4:00 271.5 F 29.8

Min 271.0 F 29.2  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 210.3 F 0.0  
10:00 199.8 F 0.0  
15:00 197.3 F 0.0  
20:00 194.3 F 0.0  
25:00 190.8 F 0.0  
30:00 187.7 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 25:46  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:35  
CYCLE COMPLETE

05/02/2024 02:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_