
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005922
12:30 06/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:04
STERILIZING 3:06
136.4 C 224kPa 3:06
136.2 C 221kPa 3:36
136.4 C 224kPa 4:06
136.4 C 223kPa 4:36
136.4 C 224kPa 5:06
136.1 C 220kPa 5:36
136.2 C 220kPa 6:06
135.9 C 222kPa 6:36
135.9 C 222kPa 6:36
VENTING 6:36
AIR DRYING 7:25
CYCLE COMPLETE 14:23

Digital Signature #
DBE99C32C0EE8C55
