

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002489  
11:57 07/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:43  
STERILIZING 3:31  
136.0 C 220kPa 3:31  
136.3 C 226kPa 4:01  
135.9 C 221kPa 4:31  
136.2 C 223kPa 5:01  
136.4 C 225kPa 5:31  
136.4 C 225kPa 6:01  
136.2 C 222kPa 6:31  
136.4 C 222kPa 7:01  
136.4 C 222kPa 7:01  
VENTING 7:01  
AIR DRYING 7:51  
CYCLE COMPLETE 37:51  
-----

Digital Signature #  
B59CE85F662C9BD0  
-----