
Midmark M11 - v1.0.5
Total Cycles: 2503

Name: RITTER2
Sterilizer ID: V2313228
06/12/2023 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 193.9 F 4.0
4:00 207.5 F 9.7
6:00 222.6 F 7.2
8:00 235.1 F 12.0
10:00 236.5 F 10.3
12:00 236.6 F 9.3
14:00 251.4 F 16.7
16:00 263.7 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.2 F 29.5
2:00 272.3 F 30.0
3:00 271.6 F 29.2
4:00 272.4 F 29.9

Min 271.3 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 229.1 F 0.0
10:00 249.9 F 0.0
15:00 230.8 F 0.0
20:00 225.8 F 0.0
25:00 227.1 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 17:56
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

06/12/2023 04:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____