
Midmark M11 - v1.0.5
Total Cycles: 4269

Name: RITTER2
Sterilizer ID: V2313228
01/16/2025 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 71.6 F 0.0
4:00 80.1 F 0.0
6:00 98.1 F 0.1
8:00 152.1 F 0.2
10:00 198.1 F 0.4
12:00 201.3 F 0.8
14:00 209.4 F 2.8
16:00 217.7 F 5.7
18:00 228.0 F 9.0
20:00 237.9 F 12.8
22:00 235.7 F 9.5
24:00 241.1 F 10.9
26:00 244.8 F 13.0
28:00 243.3 F 12.1
30:00 255.0 F 18.2
32:00 265.5 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.7 F 29.3
2:00 271.6 F 28.6
3:00 272.7 F 29.4
4:00 271.8 F 28.6

Min 271.3 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 224.3 F 0.0
10:00 234.2 F 0.0
15:00 226.0 F 0.0
20:00 229.9 F 0.0
25:00 218.2 F 0.0
30:00 212.9 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:54
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:12:34
CYCLE COMPLETE

01/16/2025 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____