
Midmark M11 - v1.0.0
Total Cycles: 4510

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2024 12:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.6 F 0.0
2:00 108.7 F 1.0
4:00 131.8 F 3.8
6:00 157.5 F 6.7
8:00 176.5 F 9.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 191F
CHAMBER PRESS 12.4PSI
** ITEMS NOT STERILE **

05/01/2024 12:47 PM