

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3039

Name: RITTER 1  
Sterilizer ID: V2178198  
03/27/2023 02:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 166.0 F 3.6  
4:00 180.0 F 7.7  
6:00 202.1 F 12.7  
8:00 219.6 F 7.1  
10:00 233.7 F 12.3  
12:00 234.5 F 9.0  
14:00 238.9 F 9.7  
16:00 246.7 F 13.7  
18:00 259.0 F 20.4  
20:00 270.1 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.3 F 28.9  
2:00 271.5 F 28.7  
3:00 272.2 F 28.9  
4:00 271.4 F 28.6

Min 271.3 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 222.2 F 0.0  
10:00 218.2 F 0.0  
15:00 220.0 F 0.0  
20:00 222.2 F 0.0  
25:00 223.9 F 0.0  
30:00 225.7 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 20:51  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

03/27/2023 03:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_