
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000494
11:05 23/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 3:57
STERILIZING 5:09
136.0 C 227kPa 5:09
136.2 C 230kPa 5:39
136.4 C 233kPa 6:09
136.4 C 232kPa 6:39
136.0 C 229kPa 7:09
136.2 C 231kPa 7:39
136.4 C 233kPa 8:09
136.4 C 232kPa 8:39
136.4 C 232kPa 8:39
VENTING 8:39
AIR DRYING 9:32
CYCLE COMPLETE 24:33

Digital Signature #
89691C49E0048584
