

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001074  
12:23 12/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:22  
PRESSURIZING 3:26  
STERILIZING 4:46  
136.0 C 220kPa 4:46  
136.4 C 225kPa 5:16  
136.4 C 222kPa 5:46  
136.3 C 224kPa 6:16  
136.3 C 225kPa 6:46  
136.4 C 224kPa 7:16  
136.4 C 225kPa 7:46  
136.4 C 224kPa 8:16  
136.4 C 224kPa 8:16  
VENTING 8:16  
AIR DRYING 9:13  
CYCLE COMPLETE 24:13  
-----

Digital Signature #  
87301B731B0DE80D  
-----