
Midmark M11 - v1.0.0
Total Cycles: 3393

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2023 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.6 F 0.0
2:00 156.8 F 0.0
4:00 168.2 F 0.5
6:00 192.8 F 4.8
8:00 209.9 F 9.6
10:00 225.5 F 8.0
12:00 231.5 F 9.5
14:00 237.2 F 9.9
16:00 242.5 F 12.2
18:00 240.2 F 10.6
20:00 253.1 F 17.0
22:00 264.3 F 23.5
24:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 214.2 F 0.0
10:00 211.4 F 0.0
15:00 218.6 F 0.0
20:00 224.8 F 0.0
25:00 225.2 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

06/15/2023 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____