

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001370  
17:15 28/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:19  
STERILIZING 4:51  
136.3 C 225kPa 4:51  
136.2 C 224kPa 5:21  
136.4 C 226kPa 5:51  
136.4 C 225kPa 6:21  
136.4 C 225kPa 6:51  
136.3 C 223kPa 7:21  
136.2 C 224kPa 7:51  
136.4 C 224kPa 8:21  
136.4 C 224kPa 8:21  
VENTING 8:21  
AIR DRYING 9:18  
CYCLE COMPLETE 21:55  
-----

Digital Signature #  
E51E44218ED1BD6E  
-----