
Midmark M11 - v1.0.0
Total Cycles: 2985

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2023 08:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.0 F 0.0
2:00 80.2 F 0.0
4:00 89.4 F 0.0
6:00 116.5 F 0.0
8:00 150.6 F 1.4
10:00 178.7 F 4.6
12:00 199.0 F 8.1
14:00 209.1 F 12.2
16:00 217.8 F 5.4
18:00 229.4 F 9.5
20:00 237.3 F 10.6
22:00 239.7 F 11.0
24:00 236.7 F 8.8
26:00 250.3 F 15.3
28:00 261.6 F 21.7
30:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 29.0
2:00 271.8 F 28.2
3:00 272.3 F 29.0
4:00 272.4 F 28.6

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 211.6 F 0.0
10:00 220.6 F 0.0
15:00 221.0 F 0.0
20:00 225.2 F 0.0
25:00 226.1 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:27
CYCLE COMPLETE

03/09/2023 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____