

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3833

Name: RITTER 1  
Sterilizer ID: V2178198  
09/28/2023 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
2:00 197.5 F 2.2  
4:00 198.3 F 6.5  
6:00 209.9 F 11.2  
8:00 221.2 F 5.6  
10:00 231.8 F 10.6  
12:00 232.5 F 7.5  
14:00 244.6 F 13.3  
16:00 243.6 F 11.9  
18:00 255.4 F 18.1  
20:00 266.0 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 28.9  
2:00 271.3 F 28.1  
3:00 272.3 F 29.0  
4:00 272.0 F 28.5

Min 271.1 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 228.1 F 0.0  
10:00 221.4 F 0.0  
15:00 221.2 F 0.0  
20:00 223.2 F 0.0  
25:00 227.6 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 21:46  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:00:50  
CYCLE COMPLETE

09/28/2023 10:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_