

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2390

Name: RITTER2  
Sterilizer ID: V2313228  
05/18/2023 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.4 F 0.0  
2:00 116.3 F 0.0  
4:00 130.8 F 0.2  
6:00 170.4 F 2.6  
8:00 192.0 F 6.4  
10:00 218.8 F 11.1  
12:00 226.9 F 5.9  
14:00 233.5 F 10.9  
16:00 233.1 F 7.8  
18:00 243.1 F 12.0  
20:00 245.2 F 12.9  
22:00 257.2 F 19.5  
24:00 268.2 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.7 F 29.4  
2:00 271.4 F 28.6  
3:00 272.0 F 29.4  
4:00 272.4 F 29.2

Min 271.0 F 28.6  
Max 273.2 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.9 F 0.0  
5:00 209.9 F 0.0  
10:00 217.2 F 0.0  
15:00 205.0 F 0.0  
20:00 204.8 F 0.0  
25:00 211.5 F 0.0  
30:00 219.0 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:16  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:03:24  
CYCLE COMPLETE

05/18/2023 03:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_