

-----  
Midmark M11 - v1.0.0  
Total Cycles: 755

Name: RITTER 1  
Sterilizer ID: V2178198  
03/29/2021 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.2 F 0.0  
2:00 180.4 F 2.8  
4:00 200.6 F 7.6  
6:00 226.4 F 10.1  
8:00 226.6 F 9.5  
10:00 234.9 F 8.4  
12:00 245.3 F 12.6  
14:00 245.4 F 12.6  
16:00 258.4 F 19.5  
18:00 269.9 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 28.5  
1:00 271.9 F 28.7  
2:00 272.5 F 28.9  
3:00 272.1 F 28.7  
4:00 272.1 F 28.5  
  
Min 271.6 F 28.4  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 247.4 F 0.0  
10:00 234.5 F 0.0  
15:00 205.3 F 0.0  
20:00 208.1 F 0.0  
25:00 178.1 F 0.0  
30:00 169.2 F 0.0  
  
Drying Complete

FILLING: 1:32  
HEATING: 19:20  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

03/29/2021 12:12 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01