
Midmark M11 - v1.0.5
Total Cycles: 3148

Name: RITTER2
Sterilizer ID: V2313228
11/20/2023 05:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 172.3 F 2.9
4:00 192.5 F 7.5
6:00 212.2 F 12.4
8:00 222.3 F 6.8
10:00 233.6 F 11.8
12:00 235.4 F 9.4
14:00 238.2 F 9.6
16:00 249.1 F 15.1
18:00 261.0 F 22.0
20:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.5
2:00 271.5 F 28.9
3:00 272.8 F 29.6
4:00 271.9 F 28.9

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 210.0 F 0.0
10:00 217.3 F 0.0
15:00 217.8 F 0.0
20:00 216.8 F 0.0
25:00 216.2 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

11/20/2023 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____