

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1584

Name: RITTER2  
Sterilizer ID: V2313228  
10/12/2022 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.3 F 0.0  
2:00 168.2 F 3.4  
4:00 175.6 F 7.3  
6:00 197.7 F 12.0  
8:00 218.7 F 6.2  
10:00 232.3 F 11.3  
12:00 232.5 F 8.5  
14:00 240.1 F 10.9  
16:00 245.1 F 13.6  
18:00 257.7 F 20.1  
20:00 268.7 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.4 F 29.9  
2:00 272.0 F 29.4  
3:00 271.1 F 29.0  
4:00 271.8 F 29.8  
  
Min 271.0 F 28.9  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 225.1 F 0.0  
10:00 235.6 F 0.0  
15:00 228.5 F 0.0  
20:00 220.3 F 0.0  
25:00 218.9 F 0.0  
30:00 216.4 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

10/12/2022 12:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_