
Midmark M11 - v1.0.0
Total Cycles: 3345

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2023 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 183.1 F 1.4
4:00 204.8 F 6.5
6:00 230.0 F 11.3
8:00 227.1 F 7.0
10:00 239.8 F 12.4
12:00 237.9 F 10.2
14:00 237.9 F 8.5
16:00 250.5 F 15.6
18:00 262.3 F 22.3
20:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.4 F 28.3
3:00 272.5 F 29.1
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 221.7 F 0.0
10:00 225.2 F 0.0
15:00 221.4 F 0.0
20:00 221.4 F 0.0
25:00 218.4 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

06/06/2023 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____