
Midmark M11 - v1.0.0
Total Cycles: 3517

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2023 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 196.0 F 2.2
4:00 197.8 F 5.8
6:00 221.2 F 10.5
8:00 227.0 F 5.3
10:00 239.8 F 10.3
12:00 233.7 F 7.9
14:00 246.1 F 13.1
16:00 243.0 F 11.5
18:00 255.3 F 17.8
20:00 266.1 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.5 F 28.8
2:00 271.7 F 28.1
3:00 271.2 F 28.2
4:00 272.4 F 28.8

Min 271.0 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 223.5 F 0.0
10:00 227.0 F 0.0
15:00 228.9 F 0.0
20:00 229.2 F 0.0
25:00 229.2 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

07/12/2023 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____