
Midmark M11 - v1.0.5
Total Cycles: 603

Name: RITTER2
Sterilizer ID: V2313228
12/06/2021 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 168.9 F 3.3
4:00 178.9 F 7.7
6:00 201.9 F 11.6
8:00 222.9 F 8.7
10:00 234.3 F 10.1
12:00 241.8 F 13.1
14:00 242.3 F 12.4
16:00 256.1 F 19.5
18:00 268.0 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.0 F 29.6
2:00 271.9 F 30.1
3:00 271.4 F 29.2
4:00 272.2 F 30.0

Min 271.1 F 29.1
Max 272.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 245.6 F 0.0
10:00 225.6 F 0.0
15:00 217.7 F 0.0
20:00 216.8 F 0.0
25:00 217.8 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:14
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

12/06/2021 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____