
Midmark M11 - v1.0.0
Total Cycles: 722

Name: RITTER 1
Sterilizer ID: V2178198
03/17/2021 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.6 F 0.0
2:00 147.1 F 1.4
4:00 157.7 F 4.1
6:00 187.3 F 8.1
8:00 216.6 F 9.3
10:00 223.3 F 6.8
12:00 235.8 F 11.8
14:00 234.6 F 8.6
16:00 241.2 F 10.4
18:00 246.7 F 13.4
20:00 259.3 F 20.2
22:00 270.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.2 F 28.2
2:00 272.7 F 28.7
3:00 271.7 F 28.2
4:00 272.3 F 28.4

Min 271.6 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 166.9 F 0.0
5:00 203.9 F 0.0
10:00 227.8 F 0.0
15:00 250.5 F 0.0
20:00 231.2 F 0.0
25:00 208.2 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

03/17/2021 09:01 AM
ACCEPTED: X
REJECTED:
Operator: U01